

RETURN AND REPAIR FORM



Customer Information

B: _____

S: _____

Company Name: Stafford Twp Police Department, Stafford, NJ

Date Received: 04/10/2020

Date given to service: 04/10/2020

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR SA-0031

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#: _____

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	
MP Labor	Labor	1.5	
6808455	Fuel Cell	1	
6808374	Geared Engine	1	
4412036	Power Supply 7110	1	

Repair Notes: Replaced defective fuel cell and geared engine. Repaired Power supply that was causing the 7110 to power off and on resetting itself.

FUEL CELL = ARNC-0216

CAL W/QC AND OPS CHECK

Service Technician: 

Date: 05/22/2020



Delivery receipt



Customer no.
150051459

Number of report
305012597

Please reference on inquiries

Date of receiving Delivery receipt date
/ / 05/26/2020

Consignee, 150051459
STAFFORD TOWNSHIP POLICE DEPT.
ATTN: DIANA SICKINGER
PO#ARSA-0031
260 E BAY AVE
MANAHAWKIN NJ 08050-3329

Customer, 150051459
STAFFORD TOWNSHIP POLICE DEPT.
ATTN: DIANA SICKINGER
PO#ARSA-0031
260 E BAY AVE
MANAHAWKIN NJ 08050-3329

Your order

Date of order:
03/13/2020
Purch.ord.no. :
20-01975

Branch text

Draeger contact person

Service Support-IDT
800-437-2437 OPT 5
SS-IDT@draeger.com

Page 1

Quantity	Description	Part no.	Equipment - No.	Serial - No.
	ALCOTEST 7110 MK III USA	8314246	1030397178	ARSA-0031
	NJ REPAIR/ARSA-0031			

Draeger Inc.
Our Tax ID: 23-1699096
3135 Quarry Road; Telford, PA 18969
An Equal Opportunity Employer M / F / V / H
Telephone 800-437-2437
<http://www.draeger.com>

Remit to:
LOCKBOX (Standard USPS)
Draeger, Inc.
PO Box 13369
Newark, New Jersey
07101-3362

Remit to:
LOCKBOX (Overnight)
FIS Lockbox Processing
Lockbox #13369
100 Grove Road
Suite E
West Deptford, NJ 08066

Remit US Wire Transfers to:
Account Name: Draeger Inc.
Account Number: 00-494-936
Transit Routing: 021001033
SWIFT: BKTRUS33
Deutsche Bank Trust Company Americas
60 Wall Street 25th Fl, New York, NY 10005

Service note

Customer no. Number of report Date of receiving
 150051459 305012597

Please reference on inquiries

Customer
 STAFFORD TOWNSHIP POLICE DEPT.
 ATTN: DETECTIVE CHRIS MULCH
 260 E BAY AVE
 MANAHAWKIN NJ 08050-3329

Consignee
 STAFFORD TOWNSHIP POLICE DEPT.
 ATTN: DETECTIVE CHRIS MULCH
 260 E BAY AVE
 MANAHAWKIN NJ 08050-3329

Your order

Date of order:
 03/13/2020
 Your reference:
 20-01975

NL-Text

Your contact person

Service Support-IDT
 800-437-2437 OPT 5
 SS-IDT@draeger.com

Page 1

Item	Quantity	Part no.	Description
			Service Order: 142708526
			NJ REPAIR/ARSA-0031
			Eq.No. : 1030397178 SN: ARSA-0031
			Material: 8314246 ALCOTEST 7110 MK III USA
			Inv.No. :
			Location:
			NJ REPAIR/ARSA-0031
	1.5	H	R001 Repair time - normal hours
	1	EA	6808374 Geared Engine, A7110/A9510
	1	EA	6808455 Sensor Alcotest B
	1	EA	4412036 Power supply, 12V, 7110 60w phihong

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Stafford Twp Police Department 260 East Bay Ave Stafford, NJ 08050	2. Contact: Christopher Mulch 3. Phone Number: 609-597-1189 x8414	4. Date: 3/2/20 5. County: Ocean
6. Alcotest Instrument Serial Number: <p style="text-align: center;">ARSA-0031</p>	7. Simulator Component Serial Number: <p style="text-align: center;">NA</p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;">NA</p>

9. Reason for Service:

I responded to Stafford Police Department to complete the semi-annual recalibration of the Alcotest 7110 #ARSA-0031. Prior to the calibration, the alcotest reset as if the power was turned off and then back on again. The power cords and connections were checked. This occurred two more times after checking the power cords before I terminated the calibration.

10. Comments:

The alcotest was placed out of service and returned to Draeger Inc for evaluation.

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE

Last Known Sequential File #: 1576

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
12. The above Instrument/Component is placed out of service pending further evaluation.
13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Dennis J. Lutz #7045

3-2-20

Name & Badge Number (Print)
Signature
Date